

ST-120 Sound Level Calibrator

IEC 60942 : 2003 Class 1/ANSI S1.40-1984

- Sound pressure levels : at 94dB and 114dB.
- For 1/2" condenser microphone calibration.
- Low battery indication.
- Harmonic waves distortion $\pm 3\%$.
- Output feedback detection

Specifications

- Output sound level: at 94dB and 114dB.
- Frequency :
1KHz $\pm 1\%$ (Class1)-(A)
- Accuracy : 94 ± 0.3 dB
114 ± 0.5 dB
- Power: 9V battery 006P or IEC6F22 or NEDA 1604
- Operation environment :
23.0°C ± 3 °C <85%RH , 850mbar-1050mbar, ≤ 2000 M



CLASS 1